W¿ZARD™

Disinfection Guide

Full Face Masks, Nasal Masks, Nasal Pillows Masks

This guide is intended for multipatient use of WiZARD full face mask, nasal mask or nasal pillows mask in a sleep lab, clinic or hospital. If the mask is used as a single user in the home, refer to the Instructions for Use for cleaning instructions.

This guide describes Apex's recommended and validated procedures for cleaning and disinfection of the mask. However, the steps for disinfection of masks vary regionally. As a result, each healthcare facility should consult its own procedures before carrying out those within this guide.

Apex validated disinfection procedures					
Mask	Replace	High level thermal disinfection	High level chemical disinfection	Validated	
	new parts	prEN ISO 15883-1	Anioxyde™ 1000	number of	
	between	158°F (70°C):100 min	Peracetic acid 0.15% for 30 minutes	cycles	
	patients	176°F (80°C):10 min			
		194°F (90°C):1 min			
		Specific European regional			
		requirements			
		199.4℉ (93℃):10 min			
Full face masks					
WiZARD 220	*	·	V	20	
Nasal masks	•				
WiZARD 210	*	V	V	20	
Nasal Pillows n	nasks	•			
WiZARD 230		V **	V	15	

^{*} Open or remove the port cap prior to disinfecting the mask.

^{**} Short tube assembly validated to 158°F (70°C).

Validated high level disinfection procedures				
	High level thermal disinfection	High level chemical disinfection		
Disassembly	Detach the short tube, unhook the buckles and unthread the headgear straps.			
	2. Remove the silicone cushion and the forehead support pad.			
Cleaning and drying	1. Clean the mask and components with a soft bristle brush and soap. Pay extra attention to all			
	crevices and cavities during cleaning.			
	2. Rinse the mask components thoroughly under running potable water.			
	3. Allow the components to air dry out of direct sunlight before high level disinfection.			
	The headgear does not require disinfection or sterilize	ation. However, it is necessary to thoroughly wash the		
	headgear before using between patients. The cleaning procedure is detailed in the instructions for use.			

Validated high level disinfection procedures				
	High level thermal disinfection	High level chemical disinfection		
Disinfection or	Using a certified hot water disinfection	1. Soak the disinfectable mask components		
sterilization and	system, soak the disinfectable mask	(excluding the headgear) in the following		
drying	components (excluding the headgear) using	solutions according to the manufacturer's		
	a temperature-time combination:	instructions:		
	prEN ISO 15883-1	Peracetic acid 0.15%(Anioxyde 1000)for		
	• 158 °F (70 °C) for 100 minutes	30 minutes		
	(Short tube of WiZARD 230 was validated to 70 °C only.)	2. Rinse the mask components in potable wate		
	• 176°F (80°C) for 10 minutes	(five litres per mask).		
	• 194°F (90°C) for 1 minute	3. Allow the mask components to air dry out of		
	Specific European regional requirements	direct sunlight.		
	• 199.4 °F (93 °C) for 10 minnutes			
	2. Once completion, remove the mask			
	components from the hot water disinfection			
	system.			
	3. Allow the mask components to air dry out of			
	direct sunlight.			
Inspection	Inspect each mask component. If any visible deterioration of mask component is apparent			
	(cracking, crazing, tears etc), the mask component should be discarded and replaced. Slight			
	discolouration of the silicone components may occur and is acceptable.			
Reassembly	Reassemble the mask according to the instructions in the Instructions for Use.			
Packaging and	Store in a dry, dust-free environment away from direct sunlight.			
storage	Storage temperature: -4°F to 140°F (-20°C to 60°C).			

Disassembling the forehead support pad before cleaning

These instructions only apply to full face mask and nasal mask.

1. Firmly push the forehead support pad against the mask frame then pull forehead support pad out of the mask frame.

Cleaning the headgear

The headgear does not require disinfection. Thoroughly washing the headgear between patients is acceptable.

- 1. Handwash the headgear in warm 86°F (30°C) water with mild soap.
- 2. Rinse well in fresh water and allow it to air dry out of direct sunlight.

- Mask components should not be subjected to autoclave or ethylene-oxide gas sterilization.
- When using detergents, disinfectants or sterilization agents, always follow the manufacturer's instructions.
- Do not iron the headgear as the material is heat sensitive and will be damaged.