







# 检验报告 TEST REPORT







Scan is following us

Scan checks for authenticity



浙江省轻工业品质量检验研究院

**Zhejiang Light Industrial Products Inspection and Research Institute** 

国家纺织服装产品质量监督检验中心(浙江)

 ${\bf National\ Textiles\ and\ Garment\ Quality\ Supervision\ Inspection\ Center}({\bf Zhejiang})$ 

### Zhejiang Light Industrial Products Inspection and Research Institute National Textiles and Garment Quality Supervision Inspection Center(Zhejiang)

### **Test report**

**Number:**W202009307E page1/2

	1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1		000+b T	antion Dood Changui N			
Name of Custome	ZheJiang LingSheng Medical	Address	889th Lantian Road, Chengxi New District, YongKang City, ZheJiang				
Traine of Gustollie	Supplies Co., Ltd		DISCIPC	Province, China.			
V C	·····································	A 1.1					
Manufacturer		Address		—————————————————————————————————————			
(江) 整 ( )	Name of sample: KN95mask	2017年日本					
Sample information	Characteristics of sample: white						
	Trademark of sample:						
	Specification/model:						
	Level: KN95						
	Category of safety specification:						
	Art. No.:						
				10.5			
<b>《大业型原</b> 版	Above-mentioned information by	Customer-	supplied	<b>一种种种种种种种种种种种种种种种种种种种种种种种种种种种种种种种种种种种种</b>			
The sent way o	f	Comple o		40 pieces			
sample	Courter	Sample q	uantity	40 pieces			
Receiving Date	of 2020/04/23	Test Ca	tegory	Entrusted inspection			
Sample		li disella					
Date of Testin	2020-04-23- 2020-04-23						
PotingPossing	GB 2626-2006						
RatingRequiremen	GD 2020-2000						
Test Summary:							
See the attached page for the results.							
See the a	tached page for the results.			2.1 北京市工业市市			
	· 签章 美						
	Test Seal						
Date of Approval: 2020-04-23							
	W 10 10 10 10 10 10 10 10 10 10 10 10 10						
Pamaris -	sample is not pretreated and the result is for reference.						
Remarks							
				and the second			
	1						

Approved by:



## **Test report**

**Number:**W202009307E page2/2

ITEM	STANDARD (KN95)	RESULT	RATING	
1. FILTER EFFICIENCY(SALT MEDIUM)	<b>建筑</b> 是最重要			
(GB 2626-2006 6.3)(%)				
- 《	≥95.0	99.8	PASS	
2. RESPIRATORY RESISTANCE				
(GB 2626-2006 6.5 & 6.6)(Pa)				
-INSPIRATORY RESISTANCE	≤350	120	PASS	
-EXPIRATORY RESISTANCE	≤250	77	PASS	
3. HEADBAND (GB 2626-2006 6.11)				
134	≥10N,hold on10s	No breakage and slippage were observed	PASS	
4. APPEARANCE (GB 2626-2006 6.1 & 6.2)		AND INCOME.		
-	Meet the requirements of GB 2626-2006 standard	The appearance quality of 2 masks conforms to the standard	PASS	
5.respiratory value cover				
· · · · · · · · · · · · · · · · · · ·	Meet the requirements of GB 2626-2006 standard	Meet the standard	PASS	

Picture(s) of sample



#### **DECLARATION**

- 1. Our organization guarantees impartiality, independence and honesty of inspection, and is responsible for the results of inspection, keeping the samples supplied by the entrusting party confidential and at the same time protecting the ownership of the samples supplied.
- 2. The test report will be deemed invalid without signatures of the inspector/reviewer and authorized personnel, and the red special inspection stamp of our organization.
- 3. The test report will be invalid if it is altered. Copies of the report are invalid without the red special inspection stamp of our organization.
- 4. The test results shown in this report are applicable only to the samples provided by customers.
- 5. All the pages of the report are integral parts of the report. Our organization will not be responsible for any misunderstanding or other results caused by using separate page(s) of the report.
- 6. If there is any dissent of the report, the entrusting party shall notify our organization timely. For the mandatory inspection given by governmental administration departments, and dissent about the sample being tested or test results on the report should be dealt with in accordance with national regulations.

### 浙江省轻工业品质量检验研究院及附设的检验中心

**The Affiliated Inspection Centers** 

浙江省轻工业品质量检验研究院

Zhejiang Light Industrial ProductsInspectionandResearchInstitute

国家纺织服装产品质量监督检验中心(浙江)

National Textiles and Garment Quality Supervision Inspection Center(Zhejiang)

地址: 浙江省杭州市江干区下沙路 300 号 6 号楼

Address: Building No.6, 300 XiaSha Road, Hangzhou, Zhejiang

国家家具产品质量监督检验中心(浙江)

National Center for Quality Supervision Inspection of Furniture(Zhejiang)

浙江省室内安全及家具产品质量检验中心

Zhejiang Center of Quality Test for Indoor Safety and Furniture Products

地址: 浙江省杭州市余杭区良渚街道经一路 1 号良渚大学科技园 4 号楼 Address: Building 4 LiangZhu University Science and Technology Park, No.1,

JingYi Rd, Yuhang District, Hangzhou, Zhejiang

浙江省轻工及五金产品质量检验中心

Zhejiang Center of Quality Test for Light Industry and Hardware Products

浙江省体育用品质量检验中心

Zhejiang Center of Quality Test for Sports Products

地址:浙江省杭州市西湖区天目山路 222 号 3 号楼

浙江省杭州市余杭区良渚街道经一路1号良渚大学科技园3号楼

Address: No. 222 Tianmushan Rd., Hangzhou, Zhejiang

Building 3 LiangZhu University Science and Technology Park, No.1,

JingYi Rd., Hangzhou, Zhejiang

国家锁具产品质量监督检验中心 (浙江)

National Center for Quality Supervision Inspection of Lock(Zhejiang)

浙江省锁具产品质量检验中心

Zhejiang Center of Quality Test for Lock Products

地址: 浙江省杭州市塘苗路 24 号

Address: No. 24 Tangmiao Rd., Hangzhou, Zhejiang

投诉电话: 0571-85023552

网上业务受理/报告查询: http://www.zjttj.cn:807 投诉电话: 0571-85023552
Online Business-Reception/Report Inquires: http://www.zjttj.cn:807 Complaint Tel: 0571-85023552

联系电话: 0571-89001107 Telephone: 0571-89001107

联系电话: 0571-85122669

Telephone: 0571-85122669

联系电话: 0571-89009556

Telephone: 0571-89009556

E-Mail: 2047699564@qq.com

E-Mail: 1137907994@qq.com

E-Mail: wujtest@126.com

联系电话: 0571-85027738

Telephone: 0571-85027738

E-Mail: Locktest@126.com